



Vitamin K1

1. Analysis Condition

YL9100 Series Vitamin Analyzer

Mobile Phase : Methanol

Flow Rate : 1.0ml/min

Column Oven : 40°C

Detector : UV/Vis 280nm

Column : C18 (4.6*250mm, 5um)

Injection Volumn : 20ul sample loop

2. Chromatogram

