

Subject : Analysis of (Diphenhydramine Hydrochloride)**► Analysis Condition****Acme 6100 Series**

Oven : 200°C (Isothermal)

Column : HP-1 (30m*0.32mm*0.25um)

Carrier gas : He, 4.0 ml/min (Split ratio 6:1)

Injector : Capillary 300°C

Detector : FID 300°C

Dilution in Ethanol → 1ul Injection

► Chromatogram