



Analysis of N-methyl Morpholine and Morpholine

Analysis Condition

Oven temp.: 80°C (10min)

Injector: Capillary, 230°C

Carrier gas: He

Flow rate: 3.0ml/min

Split ratio: 10:1

Detector: FID (250°C)

Capillary Column: HP-5 (30m x 0.32mm x 1.0um)

Sample: Morpholine , 4-Methyl Morpholine in Butanol

