



## Analysis of the Packing Material of Food

**Column:** HP-5(30m x 0.32mm x1.0um)

**Split ratio** 30:1 (Column 1ml/min, Inject 1 ml after heating at 80°C for 30min)

**Oven Program :** 35°C(8min) → 10°C/min → 50°C(1min)

**Internal Standard :** THF 0.5ul

**Sample collection area:** 8cm x 10cm (use 20ml Vial )

