



## *Analysis of Residual Solvents in Packing Materials of Food*

### **Analysis Condition**

**Column:** HP-1 (25m x 0.2mm x 0.11um)

**Detector:** FID

**Sample Preparation:** Volatilization at 80°C for 30min

**Sample:** MeOH, IPA, MEK, TOL



<Residual Solvent Analyzer>

