



Analysis of Solvent Purity(1,6-Dichlorohexane)

Injector: Capillary (230°C)

Column: Ultra-2 (25m x 0.32mm x 0.52um)

Detector: FID (250°C)

Flow: Column flow 2ml/min (Split ratio 40:1)

Carrier gas: He

Injection Volume: 0.5ul(Liquid)

Oven temp.: 70°C(1min) → 20°C/min → 250°C(1.0min)

Sample : 1,6-Dichlorohexane STD & Sample

