



## Analysis of Gas Generated in Toothpaste for Whitening

**Detector:** TCD

**Column:** Alltech CTR I (P/N 8700)

**Column flow:** 55 ml/min (Helium)

**Temp (Inj, Oven, Det):** 25°C

**TCDSEN:** 5    **Range:** 0

**Injection Volume:** 600ul

