



## *Analysis of Trace Level of CO, CO<sub>2</sub>, CH<sub>4</sub>*

### Analysis Condition

**Column:** Porapak Q (12 ft, 80/100)

**Injector:** 6 port Valve (2 cc sample loop)

**Column flow:** 27 mL/min

**Oven :** 40 °C

**Detector:** FID

**Temp :** 250 °C,

**Air:** 300mL/min,

**H<sub>2</sub>:** 30 mL/min

**Methanizer :** 350 °C

