



## Analysis of MCPD in Soy Sauce

### Analysis Condition

**Detector:** FID

**Column:** Ultra-1(25m x 0.25mm)

**Sample preparation :** Derivatization with Phenyl boric acid

**Oven temp. :**

70°C(1min) → 10°C/min → 120°C(0.5min)

→ 30°C/min → 280°C(10min)

