



## *Chlorhexidine Gluconate Solution in Detergent*

### **Analysis Condition**



**Column:** Symmetry C18 3.9 x 300 mm

**Flow rate:** 1ml/min

**Mobile phase:**

110 mmol NaClO<sub>4</sub>. 10 mmole H<sub>3</sub>PO<sub>4</sub>: 10mmole,  
10mmole NaH<sub>2</sub>PO<sub>4</sub> in water: Acetonitrile = 3:2

**Detector:** UV 254nm

**UV730D UV/VIS Detector**

